

The results set forth herein are provided by SGS North America Inc.

e-Hardcopy 2.0
Automated Report

Technical Report for

City of Cocoa Beach Water Reclamation

PFOS/PFOA Testing

SGS Job Number: FA55437

Sampling Dates: 06/27/18 - 06/28/18

Report to:

City of Cocoa Beach Water Reclamation
1600 Minutemen Causeway
Cocoa Beach, FL 32931
ckrick@cityofcocoabeach.com

ATTN: Christopher Krick

Total number of pages in report: 23



Test results contained within this data package meet the requirements of the National Environmental Laboratory Accreditation Program and/or state specific certification programs as applicable.

A handwritten signature in black ink that reads "Caitlin Brice".

Caitlin Brice, M.S.
General Manager

Client Service contact: Heather Wandrey 407-425-6700

Certifications: FL(E83510), LA(03051), KS(E-10327), IL(200063), NC(573), NJ(FL002), NY(12022), SC(96038001)
DoD ELAP(ANAB L2229), AZ(AZ0806), CA(2937), TX(T104704404), PA(68-03573), VA(460177),
AK, AR, IA, KY, MA, MS, ND, NH, NV, OK, OR, UT, WA, WV

This report shall not be reproduced, except in its entirety, without the written approval of SGS.

Test results relate only to samples analyzed.

Table of Contents

-1-

Section 1: Sample Summary	3
Section 2: Summary of Hits	4
Section 3: Sample Results	5
3.1: FA55437-1: 062718LPW/LPW	6
3.2: FA55437-2: 062718MW-3/MW-3	7
3.3: FA55437-3: 062718MW-4/MW-4	8
3.4: FA55437-4: 062718PAFB/PAFB	9
3.5: FA55437-5: 062718F/EFF-1	10
3.6: FA55437-6: 062718F/INF-2	11
Section 4: Misc. Forms	12
4.1: Chain of Custody	13
Section 5: MS Semi-volatiles - QC Data Summaries	15
5.1: Method Blank Summary	16
5.2: Blank Spike Summary	20
5.3: Matrix Spike/Matrix Spike Duplicate Summary	22

1

2

3

4

5



Sample Summary

City of Cocoa Beach Water Reclamation

Job No: FA55437

PFOS/PFOA Testing

Sample Number	Collected Date	Time By	Received	Matrix Code	Type	Client Sample ID
FA55437-1	06/27/18	10:23	CK	06/29/18	AQ Water	062718LPW/LPW
FA55437-2	06/27/18	11:48	CK	06/29/18	AQ Ground Water	062718MW-3/MW-3
FA55437-3	06/27/18	11:04	CK	06/29/18	AQ Ground Water	062718MW-4/MW-4
FA55437-4	06/27/18	13:02	CK	06/29/18	AQ Ground Water	062718PAFB/PAFB
FA55437-5	06/28/18	07:30	CK	06/29/18	AQ Water	062718F/EFF-1
FA55437-6	06/28/18	07:51	CK	06/29/18	AQ Water	062718F/INF-2

Summary of Hits

Job Number: FA55437
Account: City of Cocoa Beach Water Reclamation
Project: PFOS/PFOA Testing
Collected: 06/27/18 thru 06/28/18

Lab Sample ID	Client Sample ID	Result/ Qual	PQL	MDL	Units	Method
---------------	------------------	-----------------	-----	-----	-------	--------

FA55437-1 062718LPW/LPW

No hits reported in this sample.

FA55437-2 062718MW-3/MW-3

Perfluorooctanoic acid	0.0313	0.0077	0.0019	ug/l	EPA 537M BY ID
Perfluorooctanesulfonic acid	0.217	0.0077	0.0029	ug/l	EPA 537M BY ID

FA55437-3 062718MW-4/MW-4

Perfluorooctanoic acid	0.0286	0.0077	0.0019	ug/l	EPA 537M BY ID
Perfluorooctanesulfonic acid	0.101	0.0077	0.0029	ug/l	EPA 537M BY ID

FA55437-4 062718PAFB/PAFB

Perfluorooctanoic acid	0.0191	0.0077	0.0019	ug/l	EPA 537M BY ID
Perfluorooctanesulfonic acid	0.411	0.0077	0.0029	ug/l	EPA 537M BY ID

FA55437-5 062718F/EFF-1

Perfluorooctanoic acid	0.0234	0.0077	0.0019	ug/l	EPA 537M BY ID
Perfluorooctanesulfonic acid	0.261	0.0077	0.0029	ug/l	EPA 537M BY ID

FA55437-6 062718F/INF-2

Perfluorooctanoic acid	0.0122	0.0077	0.0019	ug/l	EPA 537M BY ID
Perfluorooctanesulfonic acid	0.165	0.0077	0.0029	ug/l	EPA 537M BY ID

Sample Results

Report of Analysis

Report of Analysis

Client Sample ID: 062718LPW/LPW		
Lab Sample ID: FA55437-1		Date Sampled: 06/27/18
Matrix: AQ - Water		Date Received: 06/29/18
Method: EPA 537M BY ID EPA 537 MOD		Percent Solids: n/a
Project: PFOS/PFOA Testing		

	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	2Q17138.D	1	07/14/18 12:30	NG	07/11/18 09:00	OP70832	S2Q296
Run #2							

	Initial Volume	Final Volume
Run #1	130 ml	1.0 ml
Run #2		

CAS No.	Compound	Result	PQL	MDL	Units	Q
---------	----------	--------	-----	-----	-------	---

PERFLUOROALKYL CARBOXYLIC ACIDS

335-67-1	Perfluorooctanoic acid	0.0019 U	0.0077	0.0019	ug/l	
----------	------------------------	----------	--------	--------	------	--

PERFLUOROALKYL SULFONATES

1763-23-1	Perfluorooctanesulfonic acid	0.0029 U	0.0077	0.0029	ug/l	
-----------	------------------------------	----------	--------	--------	------	--

CAS No.	ID Standard Recoveries	Run# 1	Run# 2	Limits
	13C8-PFOA	98%		50-150%
	13C8-PFOS	65%		50-150%

U = Not detected MDL = Method Detection Limit I = Result > = MDL but < PQL J = Estimated value
 PQL = Practical Quantitation Limit V = Indicates analyte found in associated method blank
 L = Indicates value exceeds calibration range N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID: 062718MW-3/MW-3	Date Sampled: 06/27/18
Lab Sample ID: FA55437-2	Date Received: 06/29/18
Matrix: AQ - Ground Water	Percent Solids: n/a
Method: EPA 537M BY ID IN HOUSE	
Project: PFOS/PFOA Testing	

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	2Q16915.D	1	07/11/18 01:30	NG	07/05/18 08:00	OP70766	S2Q294
Run #2							

Run #	Initial Volume	Final Volume
Run #1	130 ml	1.0 ml
Run #2		

CAS No.	Compound	Result	PQL	MDL	Units	Q
PERFLUOROALKYLCARBOXYLIC ACIDS						
335-67-1	Perfluorooctanoic acid	0.0313	0.0077	0.0019	ug/l	
PERFLUOROALKYLSULFONATES						
1763-23-1	Perfluorooctanesulfonic acid	0.217	0.0077	0.0029	ug/l	
CAS No.	ID Standard Recoveries	Run# 1	Run# 2	Limits		
	13C8-PFOA	68%		50-150%		
	13C8-PFOS	63%		50-150%		

U = Not detected MDL = Method Detection Limit I = Result > = MDL but < PQL J = Estimated value
 PQL = Practical Quantitation Limit V = Indicates analyte found in associated method blank
 L = Indicates value exceeds calibration range N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID: 062718MW-4/MW-4 Lab Sample ID: FA55437-3 Matrix: AQ - Ground Water Method: EPA 537M BY ID IN HOUSE Project: PFOS/PFOA Testing	Date Sampled: 06/27/18 Date Received: 06/29/18 Percent Solids: n/a
--	---

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	2Q16916.D	1	07/11/18 01:51	NG	07/05/18 08:00	OP70766	S2Q294
Run #2							

Run #	Initial Volume	Final Volume
Run #1	130 ml	1.0 ml
Run #2		

CAS No.	Compound	Result	PQL	MDL	Units	Q
PERFLUOROALKYL CARBOXYLIC ACIDS						
335-67-1	Perfluorooctanoic acid	0.0286	0.0077	0.0019	ug/l	
PERFLUOROALKYL SULFONATES						
1763-23-1	Perfluorooctanesulfonic acid	0.101	0.0077	0.0029	ug/l	
CAS No.	ID Standard Recoveries	Run# 1	Run# 2	Limits		
	13C8-PFOA	64%		50-150%		
	13C8-PFOS	60%		50-150%		

U = Not detected **MDL = Method Detection Limit** **I = Result > = MDL but < PQL** **J = Estimated value**
PQL = Practical Quantitation Limit **V = Indicates analyte found in associated method blank**
L = Indicates value exceeds calibration range **N = Indicates presumptive evidence of a compound**

Report of Analysis

3.4
3

Client Sample ID: 062718PAFB/PAFB Lab Sample ID: FA55437-4 Matrix: AQ - Ground Water Method: EPA 537M BY ID IN HOUSE Project: PFOS/PFOA Testing	Date Sampled: 06/27/18 Date Received: 06/29/18 Percent Solids: n/a
--	---

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	2Q16917.D	1	07/11/18 02:12	NG	07/05/18 08:00	OP70766	S2Q294
Run #2							

Run #	Initial Volume	Final Volume
Run #1	130 ml	1.0 ml
Run #2		

CAS No.	Compound	Result	PQL	MDL	Units	Q
PERFLUOROALKYL CARBOXYLIC ACIDS						
335-67-1	Perfluorooctanoic acid	0.0191	0.0077	0.0019	ug/l	
PERFLUOROALKYL SULFONATES						
1763-23-1	Perfluorooctanesulfonic acid	0.411	0.0077	0.0029	ug/l	
CAS No.	ID Standard Recoveries	Run# 1	Run# 2	Limits		
	13C8-PFOA	114%		50-150%		
	13C8-PFOS	102%		50-150%		

U = Not detected MDL = Method Detection Limit I = Result > = MDL but < PQL J = Estimated value
 PQL = Practical Quantitation Limit V = Indicates analyte found in associated method blank
 L = Indicates value exceeds calibration range N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID: 062718F/EFF-1	
Lab Sample ID: FA55437-5	Date Sampled: 06/28/18
Matrix: AQ - Water	Date Received: 06/29/18
Method: EPA 537M BY ID IN HOUSE	Percent Solids: n/a
Project: PFOS/PFOA Testing	

	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	2Q16799.D	1	07/08/18 22:36	NG	07/05/18 08:00	OP70766	S2Q292
Run #2							

	Initial Volume	Final Volume
Run #1	130 ml	1.0 ml
Run #2		

CAS No.	Compound	Result	PQL	MDL	Units	Q
---------	----------	--------	-----	-----	-------	---

PERFLUOROALKYL CARBOXYLIC ACIDS

335-67-1	Perfluorooctanoic acid	0.0234	0.0077	0.0019	ug/l	
----------	------------------------	--------	--------	--------	------	--

PERFLUOROALKYL SULFONATES

1763-23-1	Perfluorooctanesulfonic acid	0.261	0.0077	0.0029	ug/l	
-----------	------------------------------	-------	--------	--------	------	--

CAS No.	ID Standard Recoveries	Run# 1	Run# 2	Limits
	13C8-PFOA	98%		50-150%
	13C8-PFOS	92%		50-150%

U = Not detected MDL = Method Detection Limit
PQL = Practical Quantitation Limit
L = Indicates value exceeds calibration range

I = Result > = MDL but < PQL J = Estimated value
V = Indicates analyte found in associated method blank
N = Indicates presumptive evidence of a compound

Report of Analysis

3.6
3

Client Sample ID: 062718F/INF-2	
Lab Sample ID: FA55437-6	Date Sampled: 06/28/18
Matrix: AQ - Water	Date Received: 06/29/18
Method: EPA 537M BY ID IN HOUSE	Percent Solids: n/a
Project: PFOS/PFOA Testing	

	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	2Q16918.D	1	07/11/18 02:32	NG	07/05/18 08:00	OP70766	S2Q294
Run #2							

	Initial Volume	Final Volume
Run #1	130 ml	1.0 ml
Run #2		

CAS No.	Compound	Result	PQL	MDL	Units	Q
---------	----------	--------	-----	-----	-------	---

PERFLUOROALKYL CARBOXYLIC ACIDS

335-67-1	Perfluorooctanoic acid	0.0122	0.0077	0.0019	ug/l	
----------	------------------------	--------	--------	--------	------	--

PERFLUOROALKYL SULFONATES

1763-23-1	Perfluorooctanesulfonic acid	0.165	0.0077	0.0029	ug/l	
-----------	------------------------------	-------	--------	--------	------	--

CAS No.	ID Standard Recoveries	Run# 1	Run# 2	Limits
	13C8-PFOA	137%		50-150%
	13C8-PFOS	114%		50-150%

U = Not detected MDL = Method Detection Limit I = Result > = MDL but < PQL J = Estimated value
 PQL = Practical Quantitation Limit V = Indicates analyte found in associated method blank
 L = Indicates value exceeds calibration range N = Indicates presumptive evidence of a compound

Misc. Forms

Custody Documents and Other Forms

Includes the following where applicable:

- Chain of Custody



SGS North America Inc - Orlando
Chain of Custody

FA55437

SGS - ORLANDO JOB #: _____ PAGE _____ OF _____

4405 Vineland Road, Suite C-15 Orlando, FL 32811
TEL: 407-425-6700 FAX: 407-425-0707
www.sgs.com

Client / Reporting Information		Project Information										Analytical Information										Matrix Codes
Company Name: CITY OF COCOA BEACH		Project Name: PFOS/PFOA TESTING																				DW - Drinking Water
Address: 1600 MINUTEMEN CAUSEWAY		Street: 1600 MINUTEMEN CAUSEWAY																				GW - Ground Water
City: COCOA BEACH State: FL Zip: 32931		City: COCOA BEACH State: FL																				WW - Water
Project Contact: CHRIS KRICK Email: CKRICK@CITYOFCOCOABEACH.COM		Project #																				SW - Surface Water
Phone #: 321-868-3223		Fax #: 321-868-3379																				SO - Soil
Sampler(s) Name(s) (Printed)		Client Purchase Order #																				SL - Sludge
Sampler 1: CHRIS K - Sampler 2:																						OI - Oil
SGS Orlando Sample #	Field ID / Point of Collection	COLLECTION			CONTAINER INFORMATION														LAB USE ONLY			
		DATE	TIME	SAMPLED BY	MATRIX	TOTAL # OF BOTTLES	OTHER	NONE	HCl	NaOH	HN03	H2SO4	MACHINARIA	DI WATER	MEOH							
1	062718LPW/LPW	6/27/18	10:23	CK	DW	2		X								X	X					
2	062718MW-3/MW-3	6/27/18	11:48	CK	GW	2		X								X	X					
3	062718MW-4/MW-4	6/27/18	11:04	CK	GW	2		X								X	X					
4	062718PAF3/PAF3	6/27/18	13:02	CK	WW	2		X								X	X					
5	062718F/PEA-1	6/28/18	07:30	CK	WW	2		X								X	X					
6	062718R/INF-2	6/28/18	07:51	CK	WW	2		X								X	X					
Turnaround Time (Business days)		Data Deliverable Information										Comments / Remarks										
10 Day (Business) _____ Approved By: / Date: _____ 7 Day _____ 5 Day _____ 3 Day RUSH _____ 2 Day RUSH _____ 1 Day RUSH _____ Other _____		<input type="checkbox"/> COMMERCIAL "A" (RESULTS ONLY) <input type="checkbox"/> COMMERCIAL "B" (RESULTS PLUS QC) <input type="checkbox"/> REDT1 (EPA LEVEL 3) <input type="checkbox"/> FULLT1 (EPA LEVEL 4) <input type="checkbox"/> EDD'S																				
Rush T/A Data Available VIA Email or Lablink		Sample Custody must be documented below each time samples change possession, including courier delivery.																				
Relinquished by Sampler/Affiliation	Date Time:	Received By/Affiliation	Date Time:	Relinquished By/Affiliation	Date Time:	Received By/Affiliation	Date Time:	Relinquished By/Affiliation	Date Time:	Received By/Affiliation	Date Time:	Relinquished By/Affiliation	Date Time:	Received By/Affiliation	Date Time:							
<i>Chris Krick / COCB</i>	<i>6/28/18</i>	<i>James Blomen</i>	<i>6-28-18 1020</i>	<i>James Blomen</i>	<i>6-28-18 1145</i>	<i>James Blomen</i>	<i>6-28-18 1145</i>	<i>James Blomen</i>	<i>6-28-18 1145</i>	<i>James Blomen</i>	<i>6-28-18 1145</i>	<i>James Blomen</i>	<i>6-28-18 1145</i>	<i>James Blomen</i>	<i>6-28-18 1145</i>							
5		6		7		8																
Lab Use Only: Cooler Temperature (s) Celsius (corrected): <i>2.6</i>		http://www.sgs.com/en/terms-and-conditions																				

4.1
4

ORLD-SMT-0001-03-FORM-COC (1) Rev 031318



SGS Sample Receipt Summary

Job Number: FA55437

Client: CITY OF COCOA BEACH

Project: PFOS/PFOA TESTING

Date / Time Received: 6/29/2018 7:00:00 AM

Delivery Method: SGS

Airbill #'s: _____

Therm ID: IR 1;	Therm CF: 0.1;	# of Coolers: 1
Cooler Temps (Raw Measured) °C: Cooler 1: (2.5);		
Cooler Temps (Corrected) °C: Cooler 1: (2.6);		

Cooler Information

	Y	or	N
1. Custody Seals Present	<input checked="" type="checkbox"/>		<input type="checkbox"/>
2. Custody Seals Intact	<input checked="" type="checkbox"/>		<input type="checkbox"/>
3. Temp criteria achieved	<input checked="" type="checkbox"/>		<input type="checkbox"/>
4. Cooler temp verification	<u>IR Gun</u>		
5. Cooler media	<u>Ice (Bag)</u>		

Sample Information

	Y	or	N	N/A
1. Sample labels present on bottles	<input checked="" type="checkbox"/>		<input type="checkbox"/>	
2. Samples preserved properly	<input checked="" type="checkbox"/>		<input type="checkbox"/>	
3. Sufficient volume/containers recvd for analysis:	<input checked="" type="checkbox"/>		<input type="checkbox"/>	
4. Condition of sample	<u>Intact</u>			
5. Sample recvd within HT	<input checked="" type="checkbox"/>		<input type="checkbox"/>	
6. Dates/Times/IDs on COC match Sample Label	<input type="checkbox"/>		<input checked="" type="checkbox"/>	
7. VOCs have headspace	<input type="checkbox"/>		<input type="checkbox"/>	<input checked="" type="checkbox"/>
8. Bottles received for unspecified tests	<input type="checkbox"/>		<input checked="" type="checkbox"/>	
9. Compositing instructions clear	<input type="checkbox"/>		<input type="checkbox"/>	<input checked="" type="checkbox"/>
10. Voa Soil Kits/Jars received past 48hrs?	<input type="checkbox"/>		<input type="checkbox"/>	<input checked="" type="checkbox"/>
11. % Solids Jar received?	<input type="checkbox"/>		<input type="checkbox"/>	<input checked="" type="checkbox"/>
12. Residual Chlorine Present?	<input type="checkbox"/>		<input type="checkbox"/>	<input checked="" type="checkbox"/>

Trip Blank Information

	Y	or	N	N/A
1. Trip Blank present / cooler	<input type="checkbox"/>		<input type="checkbox"/>	<input checked="" type="checkbox"/>
2. Trip Blank listed on COC	<input type="checkbox"/>		<input type="checkbox"/>	<input checked="" type="checkbox"/>
	W	or	S	N/A
3. Type Of TB Received	<input type="checkbox"/>		<input type="checkbox"/>	<input checked="" type="checkbox"/>

Misc. Information

Number of Encores: 25-Gram _____ 5-Gram _____
 Test Strip Lot #s: pH 0-3 230315
 Residual Chlorine Test Strip Lot #: _____

Number of 5035 Field Kits: _____
 pH 10-12 219813A

Number of Lab Filtered Metals: _____
 Other: (Specify) _____

Comments SAMPLE #3 ID LABEL TIME READS 11:14.

SM001
Rev. Date 05/24/17

Technician: SHAYLAP

Date: 6/29/2018 7:00:00 AM

Reviewer: P,H

Date: 6/29/2018



4.1
4

MS Semi-volatiles

5

QC Data Summaries

Includes the following where applicable:

- Method Blank Summaries
- Blank Spike Summaries
- Matrix Spike and Duplicate Summaries

Method Blank Summary

Job Number: FA55437
 Account: COCBFL City of Cocoa Beach Water Reclamation
 Project: PFOS/PFOA Testing

Sample	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
OP70766-MB	2Q16774.D	1	07/08/18	NG	07/05/18	OP70766	S2Q292

The QC reported here applies to the following samples:

Method: EPA 537M BY ID

FA55437-2, FA55437-3, FA55437-4, FA55437-5, FA55437-6

CAS No.	Compound	Result	RL	MDL	Units	Q
335-67-1	Perfluorooctanoic acid	ND	0.0033	0.00083	ug/l	
1763-23-1	Perfluorooctanesulfonic acid	ND	0.0033	0.0012	ug/l	

CAS No.	ID Standard Recoveries	Limits
	13C4-PFBA	88% 30-140%
	13C5-PFPeA	92% 40-140%
	13C5-PFHxA	91% 50-150%
	13C4-PFHpA	89% 50-150%
	13C8-PFOA	92% 50-150%
	13C9-PFNA	91% 50-150%
	13C6-PFDA	97% 50-150%
	13C7-PFUnDA	78% 50-150%
	13C2-PFDoDA	69% 50-150%
	13C2-PFTeDA	68% 40-150%
	13C3-PFBS	89% 50-150%
	13C3-PFHxS	88% 50-150%
	13C8-PFOS	88% 50-150%
	13C8-FOSA	92% 30-140%
	d3-MeFOSAA	88% 50-150%
	13C2-6:2FTS	88% 50-150%
	13C2-8:2FTS	82% 50-150%

5.1.1
5

Method Blank Summary

Job Number: FA55437
 Account: COCBFL City of Cocoa Beach Water Reclamation
 Project: PFOS/PFOA Testing

Sample	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
OP70832-MB	2Q17123.D	1	07/14/18	NG	07/11/18	OP70832	S2Q296

The QC reported here applies to the following samples:

Method: EPA 537M BY ID

FA55437-1

CAS No.	Compound	Result	RL	MDL	Units	Q
335-67-1	Perfluorooctanoic acid	ND	0.0077	0.0019	ug/l	
1763-23-1	Perfluorooctanesulfonic acid	ND	0.0077	0.0029	ug/l	

CAS No.	ID Standard Recoveries	Limits
	13C4-PFBA	79% 30-140%
	13C5-PFPeA	78% 40-140%
	13C5-PFHxA	80% 50-150%
	13C4-PFHpA	82% 50-150%
	13C8-PFOA	78% 50-150%
	13C9-PFNA	77% 50-150%
	13C6-PFDA	81% 50-150%
	13C7-PFUnDA	64% 50-150%
	13C2-PFDoDA	57% 50-150%
	13C2-PFTeDA	54% 40-150%
	13C3-PFBS	81% 50-150%
	13C3-PFHxS	77% 50-150%
	13C8-PFOS	76% 50-150%
	13C8-FOSA	79% 30-140%
	d3-MeFOSAA	73% 50-150%
	13C2-4:2FTS	76% 50-150%
	13C2-6:2FTS	74% 50-150%
	13C2-8:2FTS	73% 50-150%

5.1.2
5

Instrument Blank

Job Number: FA55437
 Account: COCBFL City of Cocoa Beach Water Reclamation
 Project: PFOS/PFOA Testing

Sample	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
S2Q292-IBLK	2Q16692.D	1	07/07/18	NG	n/a	n/a	S2Q292

The QC reported here applies to the following samples:

Method: EPA 537M QSM5.1 B-15

FA55437-5

CAS No.	Compound	Result	RL	MDL	Units	Q
335-67-1	Perfluorooctanoic acid	ND	0.0077	0.0019	ug/l	
1763-23-1	Perfluorooctanesulfonic acid	ND	0.015	0.0038	ug/l	

CAS No.	ID Standard Recoveries	Limits
	13C4-PFBA	93% 50-150%
	13C5-PFPeA	95% 50-150%
	13C5-PFHxA	96% 50-150%
	13C4-PFHpA	95% 50-150%
	13C8-PFOA	96% 50-150%
	13C9-PFNA	92% 50-150%
	13C6-PFDA	97% 50-150%
	13C7-PFUnDA	92% 50-150%
	13C2-PFDoDA	92% 50-150%
	13C2-PFTeDA	92% 50-150%
	13C3-PFBS	94% 50-150%
	13C3-PFHxS	95% 50-150%
	13C8-PFOS	97% 50-150%
	13C8-FOSA	98% 50-150%
	d3-MeFOSAA	93% 50-150%
	13C2-4:2FTS	87% 50-150%
	13C2-6:2FTS	90% 50-150%
	13C2-8:2FTS	88% 50-150%

5.1.3
5

Instrument Blank

Job Number: FA55437
 Account: COCBFL City of Cocoa Beach Water Reclamation
 Project: PFOS/PFOA Testing

Sample	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
S2Q294-IBLK	2Q16896.D	1	07/10/18	NG	n/a	n/a	S2Q294

The QC reported here applies to the following samples:

Method: EPA 537M QSM5.1 B-15

FA55437-2, FA55437-3, FA55437-4, FA55437-6

CAS No.	Compound	Result	RL	MDL	Units	Q
335-67-1	Perfluorooctanoic acid	ND	0.0077	0.0019	ug/l	
1763-23-1	Perfluorooctanesulfonic acid	ND	0.015	0.0038	ug/l	

CAS No.	ID Standard Recoveries	Limits
	13C4-PFBA	90% 50-150%
	13C5-PFPeA	90% 50-150%
	13C5-PFHxA	91% 50-150%
	13C4-PFHpA	93% 50-150%
	13C8-PFOA	92% 50-150%
	13C9-PFNA	92% 50-150%
	13C6-PFDA	91% 50-150%
	13C7-PFUnDA	88% 50-150%
	13C2-PFDoDA	88% 50-150%
	13C2-PFTeDA	88% 50-150%
	13C3-PFBS	90% 50-150%
	13C3-PFHxS	90% 50-150%
	13C8-PFOS	92% 50-150%
	13C8-FOSA	97% 50-150%
	d3-MeFOSAA	90% 50-150%
	13C2-4:2FTS	83% 50-150%
	13C2-6:2FTS	86% 50-150%
	13C2-8:2FTS	82% 50-150%

5.1.4
5

Blank Spike Summary

Job Number: FA55437
 Account: COCBFL City of Cocoa Beach Water Reclamation
 Project: PFOS/PFOA Testing

Sample	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
OP70766-BS	2Q16773.D	1	07/08/18	NG	07/05/18	OP70766	S2Q292

The QC reported here applies to the following samples:

Method: EPA 537M BY ID

FA55437-2, FA55437-3, FA55437-4, FA55437-5, FA55437-6

CAS No.	Compound	Spike ug/l	BSP ug/l	BSP %	Limits
335-67-1	Perfluorooctanoic acid	0.0667	0.0759	114	74-130
1763-23-1	Perfluorooctanesulfonic acid	0.0617	0.0708	115	70-130

CAS No.	ID Standard Recoveries	BSP	Limits
	13C4-PFBA	93%	30-140%
	13C5-PFPeA	96%	40-140%
	13C5-PFHxA	96%	50-150%
	13C4-PFHpA	92%	50-150%
	13C8-PFOA	98%	50-150%
	13C9-PFNA	96%	50-150%
	13C6-PFDA	105%	50-150%
	13C7-PFUnDA	83%	50-150%
	13C2-PFDoDA	75%	50-150%
	13C2-PFTeDA	74%	40-150%
	13C3-PFBS	94%	50-150%
	13C3-PFHxS	93%	50-150%
	13C8-PFOS	92%	50-150%
	13C8-FOSA	78%	30-140%
	d3-MeFOSAA	91%	50-150%
	13C2-6:2FTS	100%	50-150%
	13C2-8:2FTS	100%	50-150%

* = Outside of Control Limits.

Blank Spike Summary

Job Number: FA55437
 Account: COCBFL City of Cocoa Beach Water Reclamation
 Project: PFOS/PFOA Testing

Sample	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
OP70832-BS	2Q17122.D	1	07/14/18	NG	07/11/18	OP70832	S2Q296

The QC reported here applies to the following samples:

Method: EPA 537M BY ID

FA55437-1

CAS No.	Compound	Spike ug/l	BSP ug/l	BSP %	Limits
335-67-1	Perfluorooctanoic acid	0.154	0.177	115	74-130
1763-23-1	Perfluorooctanesulfonic acid	0.142	0.169	119	70-130

CAS No.	ID Standard Recoveries	BSP	Limits
	13C4-PFBA	86%	30-140%
	13C5-PFPeA	84%	40-140%
	13C5-PFHxA	86%	50-150%
	13C4-PFHpA	87%	50-150%
	13C8-PFOA	83%	50-150%
	13C9-PFNA	80%	50-150%
	13C6-PFDA	84%	50-150%
	13C7-PFUnDA	70%	50-150%
	13C2-PFDoDA	64%	50-150%
	13C2-PFTeDA	59%	40-150%
	13C3-PFBS	88%	50-150%
	13C3-PFHxS	84%	50-150%
	13C8-PFOS	82%	50-150%
	13C8-FOSA	83%	30-140%
	d3-MeFOSAA	80%	50-150%
	13C2-4:2FTS	86%	50-150%
	13C2-6:2FTS	83%	50-150%
	13C2-8:2FTS	83%	50-150%

* = Outside of Control Limits.

Matrix Spike/Matrix Spike Duplicate Summary

Job Number: FA55437
 Account: COCBFL City of Cocoa Beach Water Reclamation
 Project: PFOS/PFOA Testing

Sample	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
OP70766-MS	2Q16786.D	1	07/08/18	NG	07/05/18	OP70766	S2Q292
OP70766-MSD	2Q16787.D	1	07/08/18	NG	07/05/18	OP70766	S2Q292
JC69096-1	2Q16785.D	1	07/08/18	NG	07/05/18	OP70766	S2Q292

The QC reported here applies to the following samples:

Method: EPA 537M BY ID

FA55437-2, FA55437-3, FA55437-4, FA55437-5, FA55437-6

CAS No.	Compound	JC69096-1 ug/l	Spike Q ug/l	MS ug/l	MS %	Spike ug/l	MSD ug/l	MSD %	RPD	Limits Rec/RPD
335-67-1	Perfluorooctanoic acid	0.000890J	0.0714	0.0829	115	0.0714	0.0841	116	1	74-130/30
1763-23-1	Perfluorooctanesulfonic acid	ND	0.0661	0.0817	124	0.0661	0.0831	126	2	70-130/30

CAS No.	ID Standard Recoveries	MS	MSD	JC69096-1	Limits
	13C4-PFBA	82%	79%	81%	30-140%
	13C5-PFPeA	83%	79%	82%	40-140%
	13C5-PFHxA	87%	83%	86%	50-150%
	13C4-PFHpA	82%	80%	82%	50-150%
	13C8-PFOA	87%	83%	81%	50-150%
	13C9-PFNA	82%	80%	75%	50-150%
	13C6-PFDA	86%	85%	81%	50-150%
	13C7-PFUnDA	79%	71%	66%	50-150%
	13C2-PFDoDA	73%	65%	57%	50-150%
	13C2-PFTeDA	79%	68%	63%	40-150%
	13C3-PFBS	79%	75%	77%	50-150%
	13C3-PFHxS	76%	72%	69%	50-150%
	13C8-PFOS	78%	70%	67%	50-150%
	13C8-FOSA	25%* a	60%	45%	30-140%
	d3-MeFOSAA	76%	73%	65%	50-150%
	13C2-6:2FTS	92%	88%	83%	50-150%
	13C2-8:2FTS	78%	77%	72%	50-150%

(a) Outside control limits.

* = Outside of Control Limits.

5.3.1
5

Matrix Spike/Matrix Spike Duplicate Summary

Job Number: FA55437
 Account: COCBFL City of Cocoa Beach Water Reclamation
 Project: PFOS/PFOA Testing

Sample	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
OP70832-MS	2Q17126.D	1	07/14/18	NG	07/11/18	OP70832	S2Q296
OP70832-MSD	2Q17127.D	1	07/14/18	NG	07/11/18	OP70832	S2Q296
JC68926-2	2Q17125.D	1	07/14/18	NG	07/11/18	OP70832	S2Q296

The QC reported here applies to the following samples:

Method: EPA 537M BY ID

FA55437-1

CAS No.	Compound	JC68926-2 ug/l	Spike Q ug/l	MS ug/l	MS %	Spike ug/l	MSD ug/l	MSD %	RPD	Limits Rec/RPD
335-67-1	Perfluorooctanoic acid	0.0135	0.167	0.197	110	0.167	0.217	122	10	74-130/30
1763-23-1	Perfluorooctanesulfonic acid	0.0816	0.154	0.305	145*	0.154	0.287	133*	6	70-130/30

CAS No.	ID Standard Recoveries	MS	MSD	JC68926-2	Limits
	13C4-PFBA	72%	72%		30-140%
	13C5-PFPeA	69%	68%		40-140%
	13C5-PFHxA	70%	70%		50-150%
	13C4-PFHpA	72%	71%		50-150%
	13C8-PFOA	69%	65%	70%	50-150%
	13C9-PFNA	73%	73%		50-150%
	13C6-PFDA	70%	69%		50-150%
	13C7-PFUnDA	63%	61%		50-150%
	13C2-PFDoDA	59%	57%		50-150%
	13C2-PFTeDA	60%	54%		40-150%
	13C3-PFBS	69%	68%		50-150%
	13C3-PFHxS	60%	59%		50-150%
	13C8-PFOS	65%	65%	68%	50-150%
	13C8-FOSA	43%	33%		30-140%
	d3-MeFOSAA	69%	70%		50-150%
	13C2-4:2FTS	80%	80%		50-150%
	13C2-6:2FTS	76%	74%		50-150%
	13C2-8:2FTS	73%	74%		50-150%

* = Outside of Control Limits.

5.3.2
5